# **Operating Instructions**

Medical Hand Grip "Modern", "Med", "Swing"

As hospital or nursing bed accessory Item 414, 429, 513



Thank you for purchasing a medical hand grip by flexxiparts. This high-quality product is subject to high safety standards and corresponds to the EU regulation 2017/745 (MDR). All versions are CE-marked and are produced entirely "Made in Germany".



Version: TS-PW-01-05 Date: 05.12.2025 (div. changes)

### 1) General notes



- Before taking it into operation, the hand grip must be inspected for transport damage.



- Always observe the operating instructions before taking this product into operation.



The operating instructions must be available to the charged personnel at all times. Non-observance of the operating instructions may cause health impairment of the user.



 These operating instructions are targeted at all persons who were directly charged with ordering that this product be taken into operation and operated.



- The product must be stored away from extreme moisture and heat.
- Installation may only be carried out by qualified specialist personnel.



- The product must only be installed at a trapeze bar intended for this product (round or oval tube). The permissible maximum load of the hand grip is 75.0 kg.
- The product is subject to the provisions of the relevant employers' liability insurance associations where used as work equipment.



- Standard medical device without antimicrobial effect.



 Medical device with antimicrobial effect. This promises effective protection against bacterial strains, viruses and fungi

## 2) Intended use, indications, contraindications & operation



- The product is intended for use as accessory for hospital and nursing beds, as pulling-up aid or to facilitate nursing tasks in the areas described below.
- It is used for bedridden children, adolescents or adults in the area of geriatric, nursing or home care who need help/support in sitting up.



Other areas of application must be agreed in writing with the manufacturer in advance. The product is not suitable for applications that deviate from the intended use and may cause potential hazards. This includes overloading beyond the permitted working load and use by untrained persons. Furthermore, the grab rail is not intended for use as sports equipment for rehabilitation exercises.





flexxiparts GmbH & Co. KG • Goltzstraße 22 • DE-32051 Herford Phone: +49 (0)5221 99 44 73 0 • Fax: +49 (0)5221 99 44 73 3 Date: 05.12.2025

info@flexxiparts.com • www.flexxiparts.com (div. changes)

- The hand grip must be installed to the trapeze bar with the slip protection. The tab permits adjustment in length. This adjustment must not be performed using any sharp-edged objects, since there is a danger of damaging the belt.
- When not in use the hand grip can be put over the trapeze bar. Ensure that the hand grip cannot drop unintendedly.

## 3) Cleaning and disinfection

- The product is produced of materials that ensure high chemicals resilience.
- The high number of cleaning agents and disinfectants available on the market may influence the properties of the product if the product is disinfected or cleaned with chemical products that were not expressly released by us. Therefore, we cannot assume full warranty of compatibility. (Observe the following notes in this respect.)



- According to AK-BWA, 9<sup>th</sup> expanded edition 2022, coordination of the treated object (med. hand grip), the treatment agent (decontamination agent, rinse aid) and the decontamination procedure is required before the product is subjected to the procedure for the first time.



- The following contents are classified as problematic and should not be used, since they may cause material failure in the med. hand grip and therefore pose a danger: Organic or inorganic substances with active chlorine (chlorine as such) and turpentine.



- Disinfectants recommended by us\*
   (partially combined disinfectant and cleaning agents):
   Antiseptica Kombi Flächen Desinfektion, APESIN rapid plus, BUDESIN DES
   LIQUID PRO, buraton rapid, Bacillol AF, Franko-Cid N, Incidin Plus,
   Kohrsolin FF, Lysoformin 3000, neoform MED FF, Pursept-A Xpress S,
   terralin liquid, WIBUplus Flächendesinfektion, 1+1 Wofasteril SC super
   \*For use according to manufacturer's specification. Must only be replaced by equal
   substances. There may be danger when combined with other agents. According to topical
   VAH disinfectant list.
- Use lukewarm water for cleaning. Use of non-abrasive, pH-neutralised common household cleaners is permissible.



 flexxiparts GmbH & Co. KG • Goltzstraße 22 • DE-32051 Herford
 Version: TS-PW-01-05

 Phone: +49 (0)5221 99 44 73 0 • Fax: +49 (0)5221 99 44 73 33
 Date: 05.12.2025

 info@flexxiparts.com • www.flexxiparts.com
 (div. changes)



- When used in the areas described in 2) and cleaned and disinfected regularly in **industrial bed washing systems**, the product should be replaced after 2 years.



- When used in the areas described in 2) and cleaned and disinfected regularly in hospital and nursing operation, the product should be replaced after 5 years.



- The product should be preventively replaced when plastic parts of the belt appear to have outward damage. It has proven beneficial to perform a functional check at least every quarter (e.g. in the scope of a safety-technical inspection).



- Observe that the entire product is replaced at all times, rather than only components.





The product is made of the plastics polypropylene (PP), polyvinylchloride (PVC) and polyamide (PA) and must only be disposed of according to the respective country-specific laws and regulations.

### 5) Safety notes



The product must never be turned several times around its vertical axis.
 This may cause material fatigue. The product must not be used as an aid for standing, e.g. from wheelchairs or the like, since it is merely a pulling-up aid.







 Before each use of the product, the user must visually check that it is functioning properly.





- Any serious incident that has occurred in relation to the device must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.



- This product is a medical device



 flexxiparts GmbH & Co. KG • Goltzstraße 22 • DE-32051 Herford
 Version: TS-PW-01-05

 Phone: +49 (0)5221 99 44 73 0 • Fax: +49 (0)5221 99 44 73 33
 Date: 05.12.2025

 info@flexxiparts.com • www.flexxiparts.com
 (div. changes)